

HEARTSINE™
TECHNOLOGIES

EC DECLARATION of CONFORMITY

HeartSine Technologies Limited
Canberra House
203 Airport Road West
Belfast
BT3 9ED

Declares that the medical device, in the range of External Defibrillators,
described hereafter as

The Samaritan® Public Access Defibrillator (PAD)

is in conformity with the essential requirements and provisions of the
Council Directive 93/42/EEC

and

is subject to the procedure set out in Annex II of the Directive
93/42/EEC under the supervision of Notified Body Number 0120, SGS
UK Ltd, Unit 202b Worle Parkway, Weston-super-Mare, United
Kingdom, BS22 6WA.

Belfast, 30th July 2004

.....
Mr A. G. Hofmann
President and CEO

Certificate of Test

Electromagnetic Compatibility
For a
Heartsine PAD

Date of Issue: 19th January 2004
SGS Reference: DUR41292/EMC/DK/04
Details of Sample: Model No.: No model number supplied
Unique Identifier: No unique identifier supplied

Date of Receipt: 1st December 2003
Date of Test: 16th December 2003
Test Required: Test for compliance with: -
CISPR 11: 1997 + A1: 1999 + A2: 2002
Group 1 , Class B

Conclusion: In the configuration tested the sample meets the requirements of the standards detailed above.

Issue Number: 1

This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at www.sgs.com. Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein.

This certificate refers only to the sample submitted for test.


This certificate shall not be reproduced except in full without the written approval of SGS UK Ltd.

PP



Test Engineer

D. King



Authorised Signatory

A.Reynard / F.Huggins
Technical Manager / Quality Manager

Certificate of Test

Electromagnetic Compatibility
For a
Heartsine PAD

Date of Issue: 19th January 2004

SGS Reference: DUR41292/EMC/DK/04

Details of Sample: Model No.: No model number supplied
Unique Identifier: No unique identifier supplied

Date of Receipt: 1st December 2003

Date of Test: 16th December 2003 To 23rd December 2003

Test Required: Test for compliance with: - RTCA/DO-160D:1997 (Section 21 only)

Conclusion: In the configuration tested the sample meets the requirements of the standards detailed above.

Issue Number: 1

This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at www.sgs.com. Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein.

This certificate refers only to the sample submitted for test.

This certificate shall not be reproduced except in full without the written approval of SGS UK Ltd.



Test Engineer

D. King

A. H. Reynard
Authorised Signatory

A. Reynard / F. Hoggins
Technical Manager / Quality Manager

Certificate of Test

Electromagnetic Compatibility
For a
Heartsine PAD

Date of Issue: 26th February 2004

SGS Reference: DUR41292/EMC/DK/04

Details of Sample: Model No.: No model number supplied
Unique Identifier: No unique identifier supplied

Date of Receipt: 1st December 2003

Date of Test: 16th December 2003 To 28th January 2004

Test Required: Test for compliance with: -
CISPR 11: 1997 + A1: 1999 + A2: 2002 Group 1 , Class B
EN 61000-4-2: 1995 + A1: 1998 + A2: 2000
EN 61000-4-8: 1993 + A1: 2000

Conclusion: In the configuration tested the sample meets the requirements of the standards detailed above.

Issue Number: 1

This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at www.sgs.com. Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein.

This certificate refers only to the sample submitted for test.

This certificate shall not be reproduced except in full without the written approval of SGS UK Ltd.



Test Engineer

D. King



Authorised Signatory

A. Reynard / F. Huggins
Technical Manager / Quality Manager



AMTAC

TEST CERTIFICATE

SUBMISSION INFORMATION

ALC. E : 76869 : 0104

10/03/2004

Sample Description: Samaritan PAD Automated External Defibrillator

Applicant: Heartsine Technologies
Canberra House
203 Airport Road West
Belfast BT3 9ED

TEST INFORMATION

Evaluation To: BSEN60601-2-4: 2003 & ANSI / AMMI DF 80: 2003

CONCLUSIONS: The item tested was found to comply with the relevant requirements of the above standards.

Signature:

Andy Clifford

AMTAC AUTHORISED SIGNATORY

This certificate should be read in conjunction with the attached report which has been prepared in accordance with and is subject to AMTAC's standard Terms & Conditions.

The results and conclusions only apply to the sample(s) submitted and do not guarantee the bulk of the items from which it was obtained to be of equal quality.

AMTAC LABORATORIES LIMITED – A BUREAU VERITAS COMPANY

NORMAN ROAD • BROADHEATH • ALTRINCHAM • CHESHIRE • WA14 4EP • UNITED KINGDOM

TELEPHONE: +44 (0)161 928 8924 • FAX: +44 (0)161 927 7359 • EMAIL: admin@amtac.co.uk • WEB: www.amtac.co.uk